

REPORT OF RADIOCARBON DATING ANALYSES

*Beta Analytic, Inc.
Miami, Florida*



*Consistent Accuracy . . .
... Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President

Ronald Hatfield
Christopher Patrick
Deputy Directors

February 21, 2011

Dr. William H. Doelle/Jim Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Results For Samples LCAFN10921, LCAFN13472, LCAFN13485,
LCAFN13837, LCAFN13945, LCAFN14278, LCAFN14307, LCAFN14606, LCAFN14609

Dear Dr. Doelle and Mr. Vint:

Enclosed are the radiocarbon dating results for nine samples recently sent to us. They each provided plenty of carbon for accurate measurements and all the analyses proceeded normally. As usual, the method of analysis is listed on the report with the results and calibration data is provided where applicable.

Note that one of the samples (LCAFN13945, Beta-292149) does not have a Measured Radiocarbon age and $^{13}\text{C}/^{12}\text{C}$ Ratio reported. This is because the sample was too small to do a separate $^{13}\text{C}/^{12}\text{C}$ ratio and AMS analysis. The only available $^{13}\text{C}/^{12}\text{C}$ ratio available to calculate a Conventional Radiocarbon Age was that determined on a small aliquot of graphite. Although this ratio corrects to the appropriate Conventional Radiocarbon Age, it is not reported since it includes laboratory chemical and detector induced fractionation.

As always, no students or intern researchers who would necessarily be distracted with other obligations and priorities were used in the analyses. We analyzed them with the combined attention of our entire professional staff.

If you have specific questions about the analyses, please contact us. We are always available to answer your questions.

Thank you for prepaying the analyses. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Digital signature on file

**BETA ANALYTIC INC.**

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 2/21/2011

Desert Archaeology, Incorporated

Material Received: 1/25/2011

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 292144 SAMPLE : LCAFN10921 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 3270 to 3240 (Cal BP 5220 to 5190) AND Cal BC 3110 to 2910 (Cal BP 5060 to 4860)	4410 +/- 40 BP	-25.8 o/oo	4400 +/- 40 BP
Beta - 292145 SAMPLE : LCAFN13472 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1040 to 910 (Cal BP 2990 to 2860)	2590 +/- 30 BP	-11.2 o/oo	2820 +/- 30 BP
Beta - 292146 SAMPLE : LCAFN13485 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1190 to 1140 (Cal BP 3140 to 3090) AND Cal BC 1140 to 1000 (Cal BP 3090 to 2940)	2650 +/- 30 BP	-10.4 o/oo	2890 +/- 30 BP
Beta - 292147 SAMPLE : LCAFN13837 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1120 to 970 (Cal BP 3070 to 2920) AND Cal BC 960 to 940 (Cal BP 2900 to 2890)	2630 +/- 30 BP	-10.4 o/oo	2870 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "as". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 2/21/2011

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 292149 SAMPLE : LCAFN13945 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1410 to 1210 (Cal BP 3360 to 3160) COMMENT: The original sample was too small to provide a ¹³ C/ ¹² C ratio on the original material. However, a ratio including both natural and laboratory effects was measured during the ¹⁴ C detection to calculate the true Conventional Radiocarbon Age.	NA	NA	3050 +/- 40 BP
Beta - 292150 SAMPLE : LCAFN14278 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1490 to 1360 (Cal BP 3440 to 3310) AND Cal BC 1350 to 1310 (Cal BP 3300 to 3260)	2970 +/- 40 BP	-15.5 o/oo	3130 +/- 40 BP
Beta - 292151 SAMPLE : LCAFN14307 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1190 to 1140 (Cal BP 3140 to 3090) AND Cal BC 1140 to 1000 (Cal BP 3090 to 2940)	2650 +/- 30 BP	-10.2 o/oo	2890 +/- 30 BP
Beta - 292153 SAMPLE : LCAFN14606 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 790 to 730 (Cal BP 2740 to 2680) AND Cal BC 690 to 540 (Cal BP 2640 to 2500)	2300 +/- 30 BP	-10.8 o/oo	2530 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "a". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 2/21/2011

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 292154 SAMPLE : LCAFN14609 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1880 to 1650 (Cal BP 3830 to 3600)	3170 +/- 40 BP	-8.7 ‰	3440 +/- 40 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "as". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-25.8:lab. mult=1)

Laboratory number: Beta-292144


Conventional radiocarbon age: 4400±40 BP

**2 Sigma calibrated results: Cal BC 3270 to 3240 (Cal BP 5220 to 5190) and
(95% probability) Cal BC 3110 to 2910 (Cal BP 5060 to 4860)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 3020 (Cal BP 4970)

1 Sigma calibrated result: Cal BC 3090 to 2920 (Cal BP 5040 to 4880)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.2:lab. mult=1)

Laboratory number: **Beta-292145**


Conventional radiocarbon age: **2820±30 BP**

2 Sigma calibrated result: Cal BC 1040 to 910 (Cal BP 2990 to 2860)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 980 (Cal BP 2930)

1 Sigma calibrated result: Cal BC 1010 to 920 (Cal BP 2960 to 2870)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.4;lab. mult=1)

Laboratory number: Beta-292146
Conventional radiocarbon age: 2890±30 BP
2 Sigma calibrated results: Cal BC 1190 to 1140 (Cal BP 3140 to 3090) and
(95% probability) Cal BC 1140 to 1000 (Cal BP 3090 to 2940)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1050 (Cal BP 3000)
1 Sigma calibrated result: Cal BC 1120 to 1010 (Cal BP 3070 to 2960)
(68% probability)



References:

- Database used*
INTCAL04
- Calibration Database*
INTCAL04 Radiocarbon Age Calibration
- IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).*
- Mathematics*
A Simplified Approach to Calibrating C14 Dates
- Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.4:lab. mult=1)

Laboratory number: Beta-292147


Conventional radiocarbon age: 2870±30 BP

**2 Sigma calibrated results: Cal BC 1120 to 970 (Cal BP 3070 to 2920) and
(95% probability) Cal BC 960 to 940 (Cal BP 2900 to 2890)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1020 (Cal BP 2980)

1 Sigma calibrated result: Cal BC 1080 to 1000 (Cal BP 3030 to 2950)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=N/A:lab. mult=1)

Laboratory number: Beta-292149

Conventional radiocarbon age: 3050±40 BP

2 Sigma calibrated result: Cal BC 1410 to 1210 (Cal BP 3360 to 3160)
(95% probability)

Intercept data

Intercepts of radiocarbon age

with calibration curve:

Cal BC 1360 (Cal BP 3310) and


Cal BC 1350 (Cal BP 3300) and

Cal BC 1310 (Cal BP 3260)

1 Sigma calibrated result:

Cal BC 1390 to 1270 (Cal BP 3340 to 3220)

(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-15.5;lab. mult=1)

Laboratory number: Beta-292150


Conventional radiocarbon age: 3130±40 BP

**2 Sigma calibrated results: Cal BC 1490 to 1360 (Cal BP 3440 to 3310) and
(95% probability) Cal BC 1350 to 1310 (Cal BP 3300 to 3260)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1410 (Cal BP 3360)

1 Sigma calibrated result: Cal BC 1440 to 1390 (Cal BP 3390 to 3340)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.2;lab. mult=1)

Laboratory number: Beta-292151


Conventional radiocarbon age: 2890±30 BP

2 Sigma calibrated results: Cal BC 1190 to 1140 (Cal BP 3140 to 3090) and
(95% probability) Cal BC 1140 to 1000 (Cal BP 3090 to 2940)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1050 (Cal BP 3000)

1 Sigma calibrated result: Cal BC 1120 to 1010 (Cal BP 3070 to 2960)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.8;lab. mult=1)

Laboratory number: Beta-292153


Conventional radiocarbon age: 2530±30 BP

**2 Sigma calibrated results: Cal BC 790 to 730 (Cal BP 2740 to 2680) and
(95% probability) Cal BC 690 to 540 (Cal BP 2640 to 2500)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 760 (Cal BP 2720)

**1 Sigma calibrated results: Cal BC 780 to 750 (Cal BP 2730 to 2700) and
(68% probability) Cal BC 680 to 670 (Cal BP 2630 to 2620) and
Cal BC 610 to 600 (Cal BP 2560 to 2560)**



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS


(Variables: C13/C12=-8.7;lab. mult=1)

Laboratory number: Beta-292154
Conventional radiocarbon age: 3440±40 BP
2 Sigma calibrated result: Cal BC 1880 to 1650 (Cal BP 3830 to 3600)
 (95% probability)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 1740 (Cal BP 3690)

1 Sigma calibrated results: Cal BC 1860 to 1850 (Cal BP 3810 to 3800) and
 Cal BC 1770 to 1690 (Cal BP 3720 to 3640)
 (68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com



*Consistent Accuracy . . .
... Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President
Ronald Hatfield
Christopher Patrick
Deputy Directors

October 6, 2011

Dr. William H. Doelle/J. Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Results For Samples LCAFN1455, LCAFN1673, LCAFN4610, LCAFN4981, LCAFN4985, LCAFN6981, LCAFN7593, LCAFN8185, LCAFN8965, LCAFN10144, LCAFN10181, LCAFN11137, LCAFN11178, LCAFN11548, LCAFN11909

Dear Dr. Doelle and Mr. Vint:

Enclosed are the radiocarbon dating results for 15 samples recently sent to us. They each provided plenty of carbon for accurate measurements and all the analyses proceeded normally. As usual, the method of analysis is listed on the report with the results and calibration data is provided where applicable.

As always, no students or intern researchers who would necessarily be distracted with other obligations and priorities were used in the analyses. We analyzed them with the combined attention of our entire professional staff.

If you have specific questions about the analyses, please contact us. We are always available to answer your questions.

Thank you for prepaying the analyses. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Darden Hood

Digital signature on file


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/J. Vint

Report Date: 10/6/2011

Desert Archaeology, Incorporated

Material Received: 9/28/2011

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 306683 SAMPLE : LCAFN1455 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 800 to 750 (Cal BP 2750 to 2700) AND Cal BC 690 to 660 (Cal BP 2640 to 2610) Cal BC 640 to 590 (Cal BP 2590 to 2540)	2510 +/- 30 BP	-22.3 o/oo	2550 +/- 30 BP
Beta - 306684 SAMPLE : LCAFN1673 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 760 to 680 (Cal BP 2710 to 2630) AND Cal BC 670 to 410 (Cal BP 2620 to 2360)	2210 +/- 30 BP	-10.6 o/oo	2450 +/- 30 BP
Beta - 306685 SAMPLE : LCAFN4610 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1210 to 1010 (Cal BP 3160 to 2960)	2710 +/- 30 BP	-12.6 o/oo	2910 +/- 30 BP
Beta - 306686 SAMPLE : LCAFN4981 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1010 to 890 (Cal BP 2960 to 2840) AND Cal BC 870 to 850 (Cal BP 2820 to 2800)	2550 +/- 30 BP	-10.6 o/oo	2790 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the 14C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby 14C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured 13C/12C ratios (delta 13C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta 13C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta 13C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

**BETA ANALYTIC INC.**

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/J. Vint

Report Date: 10/6/2011

Sample Data	Measured Radiocarbon Age	$^{13}\text{C}/^{12}\text{C}$ Ratio	Conventional Radiocarbon Age(*)
Beta - 306687 SAMPLE : LCAFN4985 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 920 to 810 (Cal BP 2870 to 2760)	2490 +/- 30 BP	-10.6 o/oo	2730 +/- 30 BP
Beta - 306688 SAMPLE : LCAFN6981 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1120 to 970 (Cal BP 3070 to 2920) AND Cal BC 960 to 940 (Cal BP 2900 to 2890)	2630 +/- 30 BP	-10.5 o/oo	2870 +/- 30 BP
Beta - 306689 SAMPLE : LCAFN7593 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1120 to 920 (Cal BP 3060 to 2870)	2610 +/- 30 BP	-10.4 o/oo	2850 +/- 30 BP
Beta - 306690 SAMPLE : LCAFN8185 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1260 to 1020 (Cal BP 3210 to 2980)	2920 +/- 30 BP	-24.3 o/oo	2930 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ^{14}C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ^{14}C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured $^{13}\text{C}/^{12}\text{C}$ ratios (delta ^{13}C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ^{13}C . On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ^{13}C , the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/J. Vint

Report Date: 10/6/2011

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 306691 SAMPLE : LCAFN8965 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 930 to 820 (Cal BP 2880 to 2770)	2500 +/- 30 BP	-10.3 o/oo	2740 +/- 30 BP
Beta - 306692 SAMPLE : LCAFN10144 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1000 to 840 (Cal BP 2940 to 2790)	2520 +/- 30 BP	-9.6 o/oo	2770 +/- 30 BP
Beta - 306693 SAMPLE : LCAFN10181 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 920 to 810 (Cal BP 2870 to 2760)	2500 +/- 30 BP	-10.9 o/oo	2730 +/- 30 BP
Beta - 306694 SAMPLE : LCAFN11137 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1000 to 840 (Cal BP 2940 to 2790)	2530 +/- 30 BP	-10.5 o/oo	2770 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the 14C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby 14C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured 13C/12C ratios (delta 13C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta 13C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta 13C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

**BETA ANALYTIC INC.**

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/J. Vint

Report Date: 10/6/2011

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 306695 SAMPLE : LCAFN11178 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1130 to 980 (Cal BP 3080 to 2930)	2650 +/- 30 BP	-11.0 o/oo	2880 +/- 30 BP
Beta - 306696 SAMPLE : LCAFN11548 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1300 to 1120 (Cal BP 3250 to 3060)	2930 +/- 30 BP	-22.6 o/oo	2970 +/- 30 BP
Beta - 306697 SAMPLE : LCAFN11909 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1300 to 1120 (Cal BP 3250 to 3060)	2950 +/- 30 BP	-23.6 o/oo	2970 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-22.3:lab. mult=1)

Laboratory number: Beta-306683
Conventional radiocarbon age: 2550±30 BP
2 Sigma calibrated results: Cal BC 800 to 750 (Cal BP 2750 to 2700) and
 (95% probability) Cal BC 690 to 660 (Cal BP 2640 to 2610) and
 Cal BC 640 to 590 (Cal BP 2590 to 2540)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 780 (Cal BP 2730)
 1 Sigma calibrated result: Cal BC 790 to 760 (Cal BP 2740 to 2710)
 (68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.6;lab. mult=1)

Laboratory number: Beta-306684


Conventional radiocarbon age: 2450±30 BP

**2 Sigma calibrated results: Cal BC 760 to 680 (Cal BP 2710 to 2630) and
(95% probability) Cal BC 670 to 410 (Cal BP 2620 to 2360)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 530 (Cal BP 2480)

1 Sigma calibrated results: Cal BC 740 to 690 (Cal BP 2690 to 2640) and
(68% probability) Cal BC 660 to 640 (Cal BP 2610 to 2590) and
Cal BC 550 to 490 (Cal BP 2500 to 2440) and
Cal BC 460 to 420 (Cal BP 2410 to 2370)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-12.6:lab. mult=1)

Laboratory number: Beta-306685


Conventional radiocarbon age: 2910±30 BP

**2 Sigma calibrated result: Cal BC 1210 to 1010 (Cal BP 3160 to 2960)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1120 (Cal BP 3060)

1 Sigma calibrated result: Cal BC 1130 to 1040 (Cal BP 3080 to 2990)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.6;lab. mult=1)

Laboratory number: **Beta-306686**


Conventional radiocarbon age: **2790±30 BP**

2 Sigma calibrated results: **Cal BC 1010 to 890 (Cal BP 2960 to 2840) and
(95% probability) Cal BC 870 to 850 (Cal BP 2820 to 2800)**

Intercept data

Intercept of radiocarbon age
with calibration curve: **Cal BC 920 (Cal BP 2870)**

1 Sigma calibrated result: **Cal BC 980 to 910 (Cal BP 2930 to 2860)**
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS


(Variables: C13/C12=-10.6:lab. mult=1)

Laboratory number: Beta-306687
Conventional radiocarbon age: 2730±30 BP
2 Sigma calibrated result: Cal BC 920 to 810 (Cal BP 2870 to 2760)
 (95% probability)

Intercept data

Intercepts of radiocarbon age
 with calibration curve:
 Cal BC 890 (Cal BP 2840) and
 Cal BC 870 (Cal BP 2820) and
 Cal BC 850 (Cal BP 2800)

1 Sigma calibrated result: Cal BC 910 to 830 (Cal BP 2860 to 2780)
 (68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.5;lab. mult=1)

Laboratory number: Beta-306688


Conventional radiocarbon age: 2870±30 BP

**2 Sigma calibrated results: Cal BC 1120 to 970 (Cal BP 3070 to 2920) and
(95% probability) Cal BC 960 to 940 (Cal BP 2900 to 2890)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1020 (Cal BP 2980)

1 Sigma calibrated result: Cal BC 1080 to 1000 (Cal BP 3030 to 2950)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.4:lab. mult=1)

Laboratory number: Beta-306689
Conventional radiocarbon age: 2850±30 BP
2 Sigma calibrated result: Cal BC 1120 to 920 (Cal BP 3060 to 2870)
 (95% probability)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 1010 (Cal BP 2960)
1 Sigma calibrated result: Cal BC 1040 to 980 (Cal BP 2990 to 2930)
 (68% probability)



References:

Database used
 INTCAL04
Calibration Database
 INTCAL04 Radiocarbon Age Calibration
 IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).
Mathematics
 A Simplified Approach to Calibrating C14 Dates
 Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-24.3:lab. mult=1)

Laboratory number: Beta-306690


Conventional radiocarbon age: 2930±30 BP

**2 Sigma calibrated result: Cal BC 1260 to 1020 (Cal BP 3210 to 2980)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1120 (Cal BP 3070)

**1 Sigma calibrated result: Cal BC 1200 to 1080 (Cal BP 3150 to 3030)
(68% probability)**



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.3:lab. mult=1)

Laboratory number: Beta-306691


Conventional radiocarbon age: 2740±30 BP

2 Sigma calibrated result: Cal BC 930 to 820 (Cal BP 2880 to 2770)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 900 (Cal BP 2850)

1 Sigma calibrated result: Cal BC 910 to 840 (Cal BP 2860 to 2790)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-9.6:lab. mult=1)

Laboratory number: Beta-306692


Conventional radiocarbon age: 2770±30 BP

**2 Sigma calibrated result: Cal BC 1000 to 840 (Cal BP 2940 to 2790)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 910 (Cal BP 2860)

1 Sigma calibrated result: Cal BC 930 to 900 (Cal BP 2880 to 2850)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS


(Variables: C13/C12=-10.9;lab. mult=1)

Laboratory number: Beta-306693
Conventional radiocarbon age: 2730±30 BP
2 Sigma calibrated result: Cal BC 920 to 810 (Cal BP 2870 to 2760)
 (95% probability)

Intercept data

Intercepts of radiocarbon age
 with calibration curve: Cal BC 890 (Cal BP 2840) and
 Cal BC 870 (Cal BP 2820) and
 Cal BC 850 (Cal BP 2800)

1 Sigma calibrated result: Cal BC 910 to 830 (Cal BP 2860 to 2780)
 (68% probability)



References:

Database used
 INTCAL04
Calibration Database
 INTCAL04 Radiocarbon Age Calibration
 IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).
Mathematics
 A Simplified Approach to Calibrating C14 Dates
 Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.5;lab. mult=1)

Laboratory number: Beta-306694


Conventional radiocarbon age: 2770±30 BP

**2 Sigma calibrated result: Cal BC 1000 to 840 (Cal BP 2940 to 2790)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 910 (Cal BP 2860)

1 Sigma calibrated result: Cal BC 930 to 900 (Cal BP 2880 to 2850)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11;lab. mult=1)

Laboratory number: Beta-306695
Conventional radiocarbon age: 2880±30 BP
2 Sigma calibrated result: Cal BC 1130 to 980 (Cal BP 3080 to 2930)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1040 (Cal BP 2990)
1 Sigma calibrated result: Cal BC 1120 to 1010 (Cal BP 3060 to 2960)
(68% probability)



References:

- Database used*
INTCAL04
- Calibration Database*
INTCAL04 Radiocarbon Age Calibration
IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).
- Mathematics*
A Simplified Approach to Calibrating C14 Dates
Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-22.6;lab. mult=1)

Laboratory number: Beta-306696


Conventional radiocarbon age: 2970±30 BP

**2 Sigma calibrated result: Cal BC 1300 to 1120 (Cal BP 3250 to 3060)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1210 (Cal BP 3160)

1 Sigma calibrated result: Cal BC 1260 to 1130 (Cal BP 3210 to 3080)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-23.6:lab. mult=1)

Laboratory number: Beta-306697
Conventional radiocarbon age: 2970±30 BP
2 Sigma calibrated result: Cal BC 1300 to 1120 (Cal BP 3250 to 3060)
 (95% probability)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 1210 (Cal BP 3160)
 1 Sigma calibrated result: Cal BC 1260 to 1130 (Cal BP 3210 to 3080)
 (68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com



*Consistent Accuracy . . .
... Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President
Ronald Hatfield
Christopher Patrick
Deputy Directors

October 7, 2011

Dr. William H. Doelle/James Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Results For Samples LCAFN4823, LCAFN4901, LCAFN4902, LCAFN10506, LCAFN10523, LCAFN10719, LCAFN10836, LCAFN11196, LCAFN11710, LCAFN12083, LCAFN12092, LCAFN12317, LCAFN12718, LCAFN12278, LCAFN12784, LCAFN12904, LCAFN14365

Dear Dr. Doelle and Mr. Vint:

Enclosed are the radiocarbon dating results for 17 samples recently sent to us. They each provided plenty of carbon for accurate measurements and all the analyses proceeded normally. As usual, the method of analysis is listed on the report with the results and calibration data is provided where applicable.

As always, no students or intern researchers who would necessarily be distracted with other obligations and priorities were used in the analyses. We analyzed them with the combined attention of our entire professional staff.

If you have specific questions about the analyses, please contact us. We are always available to answer your questions.

Thank you for prepaying the analyses. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Digital signature on file


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX:305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/James Vint

Report Date: 10/7/2011

Desert Archaeology, Incorporated

Material Received: 8/23/2011

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 304531 SAMPLE : LCAFN4823 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1250 to 1240 (Cal BP 3200 to 3190) AND Cal BC 1220 to 980 (Cal BP 3170 to 2930)	2660 +/- 40 BP	-10.5 o/oo	2900 +/- 40 BP
Beta - 304532 SAMPLE : LCAFN4901 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1130 to 920 (Cal BP 3080 to 2870)	2630 +/- 40 BP	-11.1 o/oo	2860 +/- 40 BP
Beta - 304533 SAMPLE : LCAFN4902 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1130 to 920 (Cal BP 3080 to 2870)	2630 +/- 40 BP	-10.7 o/oo	2860 +/- 40 BP
Beta - 304534 SAMPLE : LCAFN10506 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1250 to 1240 (Cal BP 3200 to 3190) AND Cal BC 1220 to 980 (Cal BP 3170 to 2930)	2670 +/- 40 BP	-11.1 o/oo	2900 +/- 40 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/James Vint

Report Date: 10/7/2011

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 304535 SAMPLE : LCAFN10523 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1200 to 930 (Cal BP 3150 to 2880)	2640 +/- 40 BP	-10.6 o/oo	2880 +/- 40 BP
Beta - 304536 SAMPLE : LCAFN10719 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1190 to 1140 (Cal BP 3140 to 3090) AND Cal BC 1140 to 920 (Cal BP 3090 to 2870)	2640 +/- 40 BP	-11.1 o/oo	2870 +/- 40 BP
Beta - 304537 SAMPLE : LCAFN10836 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 410 to 380 (Cal BP 2360 to 2330)	2280 +/- 30 BP	-21.3 o/oo	2340 +/- 30 BP
Beta - 304538 SAMPLE : LCAFN11196 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1260 to 1040 (Cal BP 3210 to 2990)	2690 +/- 30 BP	-9.9 o/oo	2940 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the 14C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby 14C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured 13C/12C ratios (delta 13C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta 13C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta 13C, the ratio and the Conventional Radiocarbon Age will be followed by "ass". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX:305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/James Vint

Report Date: 10/7/2011

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 304539 SAMPLE : LCAFN11710 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1200 to 930 (Cal BP 3150 to 2880)	2640 +/- 40 BP	-10.1 o/oo	2880 +/- 40 BP
Beta - 304540 SAMPLE : LCAFN12083 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1080 to 900 (Cal BP 3030 to 2850)	2550 +/- 40 BP	-8.6 o/oo	2820 +/- 40 BP
Beta - 304541 SAMPLE : LCAFN12092 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1250 to 1240 (Cal BP 3200 to 3190) AND Cal BC 1220 to 1010 (Cal BP 3170 to 2960)	2690 +/- 30 BP	-11.2 o/oo	2920 +/- 30 BP
Beta - 304542 SAMPLE : LCAFN12317 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1260 to 1020 (Cal BP 3210 to 2980)	2700 +/- 30 BP	-10.9 o/oo	2930 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/James Vint

Report Date: 10/7/2011

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 304543 SAMPLE : LCAFN12718 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1000 to 840 (Cal BP 2940 to 2790)	2740 +/- 30 BP	-23.3 o/oo	2770 +/- 30 BP
Beta - 304544 SAMPLE : LCAFN12278 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1080 to 920 (Cal BP 3030 to 2870)	2600 +/- 30 BP	-10.5 o/oo	2840 +/- 30 BP
Beta - 304545 SAMPLE : LCAFN12784 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1270 to 1050 (Cal BP 3220 to 3000)	2710 +/- 30 BP	-10.3 o/oo	2950 +/- 30 BP
Beta - 304546 SAMPLE : LCAFN12904 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1120 to 970 (Cal BP 3070 to 2920) AND Cal BC 960 to 940 (Cal BP 2900 to 2890)	2830 +/- 30 BP	-22.3 o/oo	2870 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/James Vint

Report Date: 10/7/2011

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 304547 SAMPLE : LCAFN14365 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1130 to 980 (Cal BP 3080 to 2930)	2650 +/- 30 BP	-11.0 o/oo	2880 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "****". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.5;lab. mult=1)

Laboratory number: Beta-304531


Conventional radiocarbon age: 2900±40 BP

**2 Sigma calibrated results: Cal BC 1250 to 1240 (Cal BP 3200 to 3190) and
(95% probability) Cal BC 1220 to 980 (Cal BP 3170 to 2930)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1080 (Cal BP 3030)

1 Sigma calibrated result: Cal BC 1130 to 1010 (Cal BP 3080 to 2960)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.1:lab. mult=1)

Laboratory number: Beta-304532
Conventional radiocarbon age: 2860±40 BP
2 Sigma calibrated result: Cal BC 1130 to 920 (Cal BP 3080 to 2870)
 (95% probability)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 1010 (Cal BP 2960)
1 Sigma calibrated result: Cal BC 1080 to 980 (Cal BP 3030 to 2930)
 (68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.7;lab. mult=1)

Laboratory number: **Beta-304533**


Conventional radiocarbon age: **2860±40 BP**

2 Sigma calibrated result: Cal BC 1130 to 920 (Cal BP 3080 to 2870)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1010 (Cal BP 2960)

1 Sigma calibrated result: Cal BC 1080 to 980 (Cal BP 3030 to 2930)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.1:lab. mult=1)

Laboratory number: Beta-304534


Conventional radiocarbon age: 2900±40 BP

**2 Sigma calibrated results: Cal BC 1250 to 1240 (Cal BP 3200 to 3190) and
(95% probability) Cal BC 1220 to 980 (Cal BP 3170 to 2930)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1080 (Cal BP 3030)

1 Sigma calibrated result: Cal BC 1130 to 1010 (Cal BP 3080 to 2960)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.6;lab. mult=1)

Laboratory number: **Beta-304535**


Conventional radiocarbon age: **2880±40 BP**

2 Sigma calibrated result: Cal BC 1200 to 930 (Cal BP 3150 to 2880)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1040 (Cal BP 2990)

1 Sigma calibrated result: Cal BC 1120 to 1000 (Cal BP 3070 to 2950)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.1:lab. mult=1)

Laboratory number: Beta-304536


Conventional radiocarbon age: 2870±40 BP

**2 Sigma calibrated results: Cal BC 1190 to 1140 (Cal BP 3140 to 3090) and
(95% probability) Cal BC 1140 to 920 (Cal BP 3090 to 2870)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1020 (Cal BP 2980)

1 Sigma calibrated result: Cal BC 1120 to 1000 (Cal BP 3060 to 2940)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-21.3;lab. mult=1)

Laboratory number: Beta-304537


Conventional radiocarbon age: 2340±30 BP

2 Sigma calibrated result: Cal BC 410 to 380 (Cal BP 2360 to 2330)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 400 (Cal BP 2350)

1 Sigma calibrated result: Cal BC 400 to 390 (Cal BP 2350 to 2340)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-9.9;lab. mult=1)

Laboratory number: Beta-304538


Conventional radiocarbon age: 2940±30 BP

2 Sigma calibrated result: Cal BC 1260 to 1040 (Cal BP 3210 to 2990)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1130 (Cal BP 3080)

1 Sigma calibrated result: Cal BC 1210 to 1120 (Cal BP 3160 to 3060)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.1;lab. mult=1)

Laboratory number: Beta-304539


Conventional radiocarbon age: 2880±40 BP

**2 Sigma calibrated result: Cal BC 1200 to 930 (Cal BP 3150 to 2880)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1040 (Cal BP 2990)

1 Sigma calibrated result: Cal BC 1120 to 1000 (Cal BP 3070 to 2950)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-8.6;lab. mult=1)

Laboratory number: Beta-304540


Conventional radiocarbon age: 2820±40 BP

2 Sigma calibrated result: Cal BC 1080 to 900 (Cal BP 3030 to 2850)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 980 (Cal BP 2930)

1 Sigma calibrated result: Cal BC 1010 to 920 (Cal BP 2960 to 2870)
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.2;lab. mult=1)

Laboratory number: Beta-304541


Conventional radiocarbon age: 2920±30 BP

**2 Sigma calibrated results: Cal BC 1250 to 1240 (Cal BP 3200 to 3190) and
(95% probability) Cal BC 1220 to 1010 (Cal BP 3170 to 2960)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1120 (Cal BP 3070)

**1 Sigma calibrated results: Cal BC 1190 to 1140 (Cal BP 3140 to 3090) and
(68% probability) Cal BC 1140 to 1050 (Cal BP 3090 to 3000)**



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.9;lab. mult=1)

Laboratory number: Beta-304542
Conventional radiocarbon age: 2930±30 BP
2 Sigma calibrated result: Cal BC 1260 to 1020 (Cal BP 3210 to 2980)
 (95% probability)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 1120 (Cal BP 3070)
1 Sigma calibrated result: Cal BC 1200 to 1080 (Cal BP 3150 to 3030)
 (68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.2:lab. mult=1)

Laboratory number: **Beta-304541**


Conventional radiocarbon age: **2920±30 BP**

2 Sigma calibrated results: **Cal BC 1250 to 1240 (Cal BP 3200 to 3190) and
(95% probability) Cal BC 1220 to 1010 (Cal BP 3170 to 2960)**

Intercept data

Intercept of radiocarbon age
with calibration curve: **Cal BC 1120 (Cal BP 3070)**

1 Sigma calibrated results: **Cal BC 1190 to 1140 (Cal BP 3140 to 3090) and
(68% probability) Cal BC 1140 to 1050 (Cal BP 3090 to 3000)**



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.5:lab. mult=1)

Laboratory number: Beta-304544


Conventional radiocarbon age: 2840±30 BP

2 Sigma calibrated result: Cal BC 1080 to 920 (Cal BP 3030 to 2870)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1000 (Cal BP 2950)

1 Sigma calibrated results: Cal BC 1020 to 970 (Cal BP 2980 to 2920) and
(68% probability) Cal BC 960 to 940 (Cal BP 2900 to 2890)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.3;lab. mult=1)

Laboratory number: **Beta-304545**

Conventional radiocarbon age: **2950±30 BP**


2 Sigma calibrated result: Cal BC 1270 to 1050 (Cal BP 3220 to 3000)
(95% probability)

Intercept data

Intercepts of radiocarbon age
with calibration curve:

Cal BC 1190 (Cal BP 3140) and
Cal BC 1140 (Cal BP 3090) and
Cal BC 1140 (Cal BP 3090)

1 Sigma calibrated results: Cal BC 1250 to 1240 (Cal BP 3200 to 3190) and
(68% probability) **Cal BC 1220 to 1120 (Cal BP 3170 to 3070)**



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-22.3;lab. mult=1)

Laboratory number: **Beta-304546**


Conventional radiocarbon age: **2870±30 BP**

2 Sigma calibrated results: **Cal BC 1120 to 970 (Cal BP 3070 to 2920) and
(95% probability) Cal BC 960 to 940 (Cal BP 2900 to 2890)**

Intercept data

Intercept of radiocarbon age
with calibration curve: **Cal BC 1020 (Cal BP 2980)**

1 Sigma calibrated result: **Cal BC 1080 to 1000 (Cal BP 3030 to 2950)**
(68% probability)



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11:lab. mult=1)

Laboratory number: Beta-304547


Conventional radiocarbon age: 2880±30 BP

**2 Sigma calibrated result: Cal BC 1130 to 980 (Cal BP 3080 to 2930)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1040 (Cal BP 2990)

**1 Sigma calibrated result: Cal BC 1120 to 1010 (Cal BP 3060 to 2960)
(68% probability)**



References:

Database used

INTCAL04

Calibration Database

INTCAL04 Radiocarbon Age Calibration

IntCal04: Calibration Issue of Radiocarbon (Volume 46, nr 3, 2004).

Mathematics

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2), p317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com



*Consistent Accuracy . . .
... Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President

Ronald Hatfield
Christopher Patrick
Deputy Directors

July 20, 2012

Dr. William H. Doelle/J. Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Results For Samples LCAFN349, LCAFN3973, LCAFN4397, LCAFN510,
LCAFN9099, LCAFN10639, LCAFN11092, LCAFN12344, LCAFN12347, LCAFN12953,
LCAFN13594, LCAFN14001, LCAFN14002

Dear Dr. Doelle:

Enclosed are the radiocarbon dating results for 13 samples recently sent to us. They each provided plenty of carbon for accurate measurements and all the analyses proceeded normally. As usual, the method of analysis is listed on the report with the results and calibration data is provided where applicable.

As always, no students or intern researchers who would necessarily be distracted with other obligations and priorities were used in the analyses. We analyzed them with the combined attention of our entire professional staff.

If you have specific questions about the analyses, please contact us. We are always available to answer your questions.

Thank you for prepaying the analyses. Thank you. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Darden Hood

Digital signature on file


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/J. Vint

Report Date: 7/20/2012

Desert Archaeology, Incorporated

Material Received: 7/12/2012

Sample Data	Measured Radiocarbon Age	$^{13}\text{C}/^{12}\text{C}$ Ratio	Conventional Radiocarbon Age(*)
Beta - 325653 SAMPLE : LCAFN349 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 840 to 800 (Cal BP 2790 to 2740)	2640 +/- 30 BP	-23.5 o/oo	2660 +/- 30 BP
Beta - 325654 SAMPLE : LCAFN3973 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1010 to 890 (Cal BP 2960 to 2840) AND Cal BC 880 to 850 (Cal BP 2820 to 2800)	2550 +/- 30 BP	-10.4 o/oo	2790 +/- 30 BP
Beta - 325655 SAMPLE : LCAFN4397 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1260 to 1240 (Cal BP 3200 to 3190) AND Cal BC 1210 to 1010 (Cal BP 3160 to 2960)	2690 +/- 30 BP	-11.2 o/oo	2920 +/- 30 BP
Beta - 325656 SAMPLE : LCAFN510 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 920 to 810 (Cal BP 2870 to 2760)	2470 +/- 30 BP	-9.8 o/oo	2720 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ^{14}C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ^{14}C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured $^{13}\text{C}/^{12}\text{C}$ ratios (delta ^{13}C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ^{13}C . On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ^{13}C , the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/J. Vint

Report Date: 7/20/2012

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 325657 SAMPLE : LCAFN9099 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1190 to 1180 (Cal BP 3140 to 3130) AND Cal BC 1160 to 1140 (Cal BP 3110 to 3090) Cal BC 1130 to 1000 (Cal BP 3080 to 2950)	2670 +/- 30 BP	-11.4 o/oo	2890 +/- 30 BP
Beta - 325658 SAMPLE : LCAFN10639 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1000 to 840 (Cal BP 2950 to 2790)	2550 +/- 30 BP	-10.7 o/oo	2780 +/- 30 BP
Beta - 325659 SAMPLE : LCAFN11092 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1210 to 1200 (Cal BP 3160 to 3150) AND Cal BC 1190 to 1140 (Cal BP 3140 to 3090) Cal BC 1130 to 1000 (Cal BP 3080 to 2950)	2660 +/- 30 BP	-10.2 o/oo	2900 +/- 30 BP
Beta - 325660 SAMPLE : LCAFN12344 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1120 to 970 (Cal BP 3070 to 2920) AND Cal BC 960 to 930 (Cal BP 2910 to 2880)	2850 +/- 30 BP	-24.5 o/oo	2860 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the 14C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby 14C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured 13C/12C ratios (delta 13C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta 13C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta 13C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX:305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/J. Vint

Report Date: 7/20/2012

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 325661 SAMPLE : LCAFN12347 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1190 to 1180 (Cal BP 3140 to 3130) AND Cal BC 1150 to 1150 (Cal BP 3100 to 3100) Cal BC 1130 to 980 (Cal BP 3080 to 2920)	2640 +/- 30 BP	-10.2 o/oo	2880 +/- 30 BP
Beta - 325662 SAMPLE : LCAFN12953 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1380 to 1340 (Cal BP 3330 to 3280) AND Cal BC 1320 to 1190 (Cal BP 3270 to 3140) Cal BC 1180 to 1160 (Cal BP 3130 to 3110) AND Cal BC 1140 to 1130 (Cal BP 3090 to 3080)	2770 +/- 30 BP	-10.5 o/oo	3010 +/- 30 BP
Beta - 325663 SAMPLE : LCAFN13594 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1260 to 1230 (Cal BP 3210 to 3180) AND Cal BC 1220 to 1020 (Cal BP 3170 to 2970)	2700 +/- 30 BP	-11.0 o/oo	2930 +/- 30 BP
Beta - 325664 SAMPLE : LCAFN14001 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1260 to 1230 (Cal BP 3210 to 3180) AND Cal BC 1220 to 1020 (Cal BP 3170 to 2970)	2700 +/- 30 BP	-10.8 o/oo	2930 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/J. Vint

Report Date: 7/20/2012

Sample Data	Measured Radiocarbon Age	$^{13}\text{C}/^{12}\text{C}$ Ratio	Conventional Radiocarbon Age(*)
Beta - 325665 SAMPLE : LCAFN14002 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1120 to 920 (Cal BP 3070 to 2870)	2800 +/- 30 BP	-22.2 o/oo	2850 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ^{14}C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ^{14}C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured $^{13}\text{C}/^{12}\text{C}$ ratios (delta ^{13}C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ^{13}C . On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ^{13}C , the ratio and the Conventional Radiocarbon Age will be followed by "ass". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-23.5;lab. mult=1)

Laboratory number: Beta-325653


Conventional radiocarbon age: 2660±30 BP

**2 Sigma calibrated result: Cal BC 840 to 800 (Cal BP 2790 to 2740)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 810 (Cal BP 2760)

1 Sigma calibrated result: Cal BC 830 to 800 (Cal BP 2780 to 2750)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.4;lab. mult=1)

Laboratory number: Beta-325654


Conventional radiocarbon age: 2790±30 BP

2 Sigma calibrated results: Cal BC 1010 to 890 (Cal BP 2960 to 2840) and
(95% probability) Cal BC 880 to 850 (Cal BP 2820 to 2800)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 920 (Cal BP 2870)

1 Sigma calibrated result: Cal BC 980 to 900 (Cal BP 2920 to 2850)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,
Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.2;lab. mult=1)

Laboratory number: **Beta-325655**


Conventional radiocarbon age: **2920±30 BP**

2 Sigma calibrated results: Cal BC 1260 to 1240 (Cal BP 3200 to 3190) and
(95% probability) Cal BC 1210 to 1010 (Cal BP 3160 to 2960)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1120 (Cal BP 3070)

1 Sigma calibrated results: Cal BC 1190 to 1180 (Cal BP 3140 to 3130) and
(68% probability) Cal BC 1160 to 1140 (Cal BP 3110 to 3090) and
Cal BC 1130 to 1050 (Cal BP 3080 to 3000)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS


(Variables: C13/C12=-9.8:lab. mult=1)

Laboratory number: Beta-325656
Conventional radiocarbon age: 2720±30 BP
2 Sigma calibrated result: Cal BC 920 to 810 (Cal BP 2870 to 2760)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 840 (Cal BP 2790)

1 Sigma calibrated result: Cal BC 900 to 830 (Cal BP 2850 to 2780)
(68% probability)



References:

Database used
INTCAL09

References to INTCAL09 database
Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,
Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario
A Simplified Approach to Calibrating C14 Dates
Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.4;lab. mult=1)

Laboratory number: Beta-325657


Conventional radiocarbon age: 2890±30 BP

**2 Sigma calibrated results: Cal BC 1190 to 1180 (Cal BP 3140 to 3130) and
(95% probability) Cal BC 1160 to 1140 (Cal BP 3110 to 3090) and
Cal BC 1130 to 1000 (Cal BP 3080 to 2950)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1050 (Cal BP 3000)

1 Sigma calibrated result: Cal BC 1120 to 1010 (Cal BP 3070 to 2960)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.7;lab.mult=1)

Laboratory number: Beta-325658
Conventional radiocarbon age: 2780±30 BP
2 Sigma calibrated result: Cal BC 1000 to 840 (Cal BP 2950 to 2790)
(95% probability)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 920 (Cal BP 2870)
 1 Sigma calibrated results: Cal BC 970 to 960 (Cal BP 2920 to 2910) and
 (68% probability) Cal BC 940 to 900 (Cal BP 2890 to 2850)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150, Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.2:lab. mult=1)

Laboratory number: Beta-325659

Conventional radiocarbon age: 2900±30 BP


**2 Sigma calibrated results: Cal BC 1210 to 1200 (Cal BP 3160 to 3150) and
(95% probability) Cal BC 1190 to 1140 (Cal BP 3140 to 3090) and
Cal BC 1130 to 1000 (Cal BP 3080 to 2950)**

Intercept data

Intercepts of radiocarbon age
with calibration curve:

Cal BC 1110 (Cal BP 3060) and
Cal BC 1100 (Cal BP 3050) and
Cal BC 1080 (Cal BP 3030) and
Cal BC 1060 (Cal BP 3010) and
Cal BC 1060 (Cal BP 3000)

1 Sigma calibrated result: Cal BC 1130 to 1020 (Cal BP 3080 to 2970)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,
Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-24.5:lab. mult=1)

Laboratory number: Beta-325660


Conventional radiocarbon age: 2860±30 BP

**2 Sigma calibrated results: Cal BC 1120 to 970 (Cal BP 3070 to 2920) and
(95% probability) Cal BC 960 to 930 (Cal BP 2910 to 2880)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1010 (Cal BP 2960)

1 Sigma calibrated result: Cal BC 1050 to 1000 (Cal BP 3000 to 2950)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150, Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.2;lab. mult=1)

Laboratory number: Beta-325661


Conventional radiocarbon age: 2880±30 BP

**2 Sigma calibrated results: Cal BC 1190 to 1180 (Cal BP 3140 to 3130) and
(95% probability) Cal BC 1150 to 1150 (Cal BP 3100 to 3100) and
Cal BC 1130 to 980 (Cal BP 3080 to 2920)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1050 (Cal BP 3000)

1 Sigma calibrated result: Cal BC 1120 to 1010 (Cal BP 3070 to 2960)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.5;lab. mult=1)

Laboratory number: **Beta-325662**


Conventional radiocarbon age: **3010±30 BP**

2 Sigma calibrated results: Cal BC 1380 to 1340 (Cal BP 3330 to 3280) and
(95% probability) Cal BC 1320 to 1190 (Cal BP 3270 to 3140) and
Cal BC 1180 to 1160 (Cal BP 3130 to 3110) and
Cal BC 1140 to 1130 (Cal BP 3090 to 3080)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1260 (Cal BP 3210)

1 Sigma calibrated results: Cal BC 1310 to 1260 (Cal BP 3260 to 3200) and
(68% probability) Cal BC 1240 to 1210 (Cal BP 3190 to 3160)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150, Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11:lab. mult=1)

Laboratory number: **Beta-325663**


Conventional radiocarbon age: **2930±30 BP**

**2 Sigma calibrated results: Cal BC 1260 to 1230 (Cal BP 3210 to 3180) and
(95% probability) Cal BC 1220 to 1020 (Cal BP 3170 to 2970)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1130 (Cal BP 3080)

**1 Sigma calibrated results: Cal BC 1210 to 1200 (Cal BP 3160 to 3150) and
(68% probability) Cal BC 1190 to 1140 (Cal BP 3140 to 3090) and
Cal BC 1130 to 1110 (Cal BP 3080 to 3060) and
Cal BC 1100 to 1080 (Cal BP 3050 to 3030) and
Cal BC 1060 to 1060 (Cal BP 3010 to 3000)**



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.8;lab. mult=1)

Laboratory number: Beta-325664


Conventional radiocarbon age: 2930±30 BP

**2 Sigma calibrated results: Cal BC 1260 to 1230 (Cal BP 3210 to 3180) and
(95% probability) Cal BC 1220 to 1020 (Cal BP 3170 to 2970)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1130 (Cal BP 3080)

**1 Sigma calibrated results: Cal BC 1210 to 1200 (Cal BP 3160 to 3150) and
(68% probability) Cal BC 1190 to 1140 (Cal BP 3140 to 3090) and
Cal BC 1130 to 1110 (Cal BP 3080 to 3060) and
Cal BC 1100 to 1080 (Cal BP 3050 to 3030) and
Cal BC 1060 to 1060 (Cal BP 3010 to 3000)**



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150, Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-22.2;lab. mult=1)

Laboratory number: Beta-325665


Conventional radiocarbon age: 2850±30 BP

**2 Sigma calibrated result: Cal BC 1120 to 920 (Cal BP 3070 to 2870)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1010 (Cal BP 2960)

1 Sigma calibrated result: Cal BC 1050 to 980 (Cal BP 3000 to 2920)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com



*Consistent Accuracy . . .
... Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President
Ronald Hatfield
Christopher Patrick
Deputy Directors

October 4, 2012

Dr. William H. Doelle/Jim Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Result For Sample LCAFN13472

Dear Dr. Doelle and Mr. Vint:

Enclosed is the radiocarbon dating result for one sample recently sent to us. It provided plenty of carbon for an accurate measurement and the analysis proceeded normally. As usual, the method of analysis is listed on the report sheet and calibration data is provided where applicable.

As always, no students or intern researchers who would necessarily be distracted with other obligations and priorities were used in the analysis. It was analyzed with the combined attention of our entire professional staff.

If you have specific questions about the analyses, please contact us. We are always available to answer your questions.

Thank you for prepaying the analysis. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Digital signature on file

**BETA ANALYTIC INC.**

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 10/4/2012

Desert Archaeology, Incorporated

Material Received: 9/27/2012

Sample Data	Measured Radiocarbon Age	$^{13}\text{C}/^{12}\text{C}$ Ratio	Conventional Radiocarbon Age(*)
Beta - 331700 SAMPLE : LCAFN13472 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1740 to 1710 (Cal BP 3690 to 3660) AND Cal BC 1700 to 1600 (Cal BP 3640 to 3560) Cal BC 1570 to 1560 (Cal BP 3520 to 3510) AND Cal BC 1550 to 1540 (Cal BP 3500 to 3490)	3130 +/- 30 BP	-10.7 o/oo	3360 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ^{14}C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ^{14}C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured $^{13}\text{C}/^{12}\text{C}$ ratios ($\delta^{13}\text{C}$) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the $\delta^{13}\text{C}$. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed $\delta^{13}\text{C}$, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.7:lab. mult=1)


Laboratory number: Beta-331700
Conventional radiocarbon age: 3360±30 BP
2 Sigma calibrated results: Cal BC 1740 to 1710 (Cal BP 3690 to 3660) and
 (95% probability) Cal BC 1700 to 1600 (Cal BP 3640 to 3560) and
 Cal BC 1570 to 1560 (Cal BP 3520 to 3510) and
 Cal BC 1550 to 1540 (Cal BP 3500 to 3490)

Intercept data

Intercepts of radiocarbon age
 with calibration curve:

Cal BC 1660 (Cal BP 3610) and
 Cal BC 1650 (Cal BP 3600) and
 Cal BC 1640 (Cal BP 3590)

1 Sigma calibrated result: Cal BC 1690 to 1620 (Cal BP 3640 to 3570)
 (68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,
 Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com



*Consistent Accuracy . . .
... Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President

Ronald Hatfield
Christopher Patrick
Deputy Directors

November 5, 2012

Dr. William H. Doelle/Mr. Jim Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Results For Samples LCAFN1232, LCAFN2287, LCAFN2665, LCAFN6676, LCAFN6711, LCAFN7173, LCAFN7208, LCAFN7278, LCAFN7885, LCAFN8001, LCAFN10527, LCAFN12782

Dear Mr. Vint:

Enclosed are the radiocarbon dating results for 12 samples recently sent to us. They each provided plenty of carbon for accurate measurements and all the analyses proceeded normally. As usual, the method of analysis is listed on the report with the results and calibration data is provided where applicable.

As always, no students or intern researchers who would necessarily be distracted with other obligations and priorities were used in the analyses. We analyzed them with the combined attention of our entire professional staff.

If you have specific questions about the analyses, please contact us. We are always available to answer your questions.

Thank you for prepaying the analyses. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Darden Hood

Digital signature on file


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX:305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 11/5/2012

Desert Archaeology, Incorporated

Material Received: 10/30/2012

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 333929 SAMPLE : LCAFN1232 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 910 to 810 (Cal BP 2860 to 2760)	2740 +/- 30 BP	-27.0 o/oo	2710 +/- 30 BP
Beta - 333930 SAMPLE : LCAFN2287 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 790 to 730 (Cal BP 2740 to 2680) AND Cal BC 690 to 660 (Cal BP 2640 to 2610) Cal BC 650 to 540 (Cal BP 2600 to 2490)	2280 +/- 30 BP	-10.0 o/oo	2530 +/- 30 BP
Beta - 333931 SAMPLE : LCAFN2665 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 3510 to 3420 (Cal BP 5460 to 5370) AND Cal BC 3420 to 3410 (Cal BP 5360 to 5360) Cal BC 3400 to 3400 (Cal BP 5350 to 5350) AND Cal BC 3380 to 3360 (Cal BP 5330 to 5310)	4400 +/- 30 BP	-10.3 o/oo	4640 +/- 30 BP
Beta - 333932 SAMPLE : LCAFN6676 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 830 to 790 (Cal BP 2780 to 2740)	2380 +/- 30 BP	-9.9 o/oo	2630 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 11/5/2012

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 333933 SAMPLE : LCAFN6711 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 830 to 790 (Cal BP 2780 to 2740)	2430 +/- 30 BP	-11.9 o/oo	2640 +/- 30 BP
Beta - 333934 SAMPLE : LCAFN7173 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 920 to 810 (Cal BP 2870 to 2760)	2460 +/- 30 BP	-8.7 o/oo	2730 +/- 30 BP
Beta - 333935 SAMPLE : LCAFN7208 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 520 to 390 (Cal BP 2470 to 2340)	2350 +/- 30 BP	-23.0 o/oo	2380 +/- 30 BP
Beta - 333936 SAMPLE : LCAFN7278 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 840 to 800 (Cal BP 2790 to 2740)	2430 +/- 30 BP	-11.2 o/oo	2660 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "****". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX:305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 11/5/2012

Sample Data	Measured Radiocarbon Age	$^{13}\text{C}/^{12}\text{C}$ Ratio	Conventional Radiocarbon Age(*)
Beta - 333937 SAMPLE : LCAFN7885 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 920 to 810 (Cal BP 2870 to 2760)	2480 +/- 30 BP	-10.2 o/oo	2720 +/- 30 BP
Beta - 333938 SAMPLE : LCAFN8001 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 900 to 800 (Cal BP 2850 to 2750)	2420 +/- 30 BP	-7.9 o/oo	2700 +/- 30 BP
Beta - 333939 SAMPLE : LCAFN10527 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1310 to 1120 (Cal BP 3260 to 3070)	2750 +/- 30 BP	-10.8 o/oo	2980 +/- 30 BP
Beta - 333940 SAMPLE : LCAFN12782 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 1260 to 1240 (Cal BP 3200 to 3190) AND Cal BC 1210 to 1010 (Cal BP 3160 to 2960)	2680 +/- 30 BP	-10.1 o/oo	2920 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ^{14}C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ^{14}C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured $^{13}\text{C}/^{12}\text{C}$ ratios (delta ^{13}C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ^{13}C . On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ^{13}C , the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-27:lab. mult=1)

Laboratory number: **Beta-333929**


Conventional radiocarbon age: **2710±30 BP**

2 Sigma calibrated result: Cal BC 910 to 810 (Cal BP 2860 to 2760)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 840 (Cal BP 2780)

1 Sigma calibrated result: Cal BC 900 to 820 (Cal BP 2850 to 2770)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150.

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS


(Variables: C13/C12=-10;lab. mult=1)

Laboratory number: Beta-333930
Conventional radiocarbon age: 2530±30 BP
2 Sigma calibrated results: Cal BC 790 to 730 (Cal BP 2740 to 2680) and
 (95% probability) Cal BC 690 to 660 (Cal BP 2640 to 2610) and
 Cal BC 650 to 540 (Cal BP 2600 to 2490)

Intercept data

Intercepts of radiocarbon age
 with calibration curve:
 Cal BC 760 (Cal BP 2720) and
 Cal BC 680 (Cal BP 2630) and
 Cal BC 670 (Cal BP 2620)

1 Sigma calibrated results:
 (68% probability) Cal BC 780 to 750 (Cal BP 2730 to 2700) and
 Cal BC 680 to 670 (Cal BP 2630 to 2620) and
 Cal BC 610 to 600 (Cal BP 2560 to 2550)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,
 Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.3;lab. mult=1)

Laboratory number: Beta-333931


Conventional radiocarbon age: 4640±30 BP

**2 Sigma calibrated results: Cal BC 3510 to 3420 (Cal BP 5460 to 5370) and
(95% probability) Cal BC 3420 to 3410 (Cal BP 5360 to 5360) and
Cal BC 3400 to 3400 (Cal BP 5350 to 5350) and
Cal BC 3380 to 3360 (Cal BP 5330 to 5310)**

Intercept data

Intercepts of radiocarbon age
with calibration curve: Cal BC 3490 (Cal BP 5440) and
Cal BC 3470 (Cal BP 5420) and
Cal BC 3370 (Cal BP 5320)

1 Sigma calibrated results: Cal BC 3500 to 3450 (Cal BP 5450 to 5400) and
(68% probability) Cal BC 3440 to 3440 (Cal BP 5390 to 5390) and
Cal BC 3380 to 3360 (Cal BP 5330 to 5320)



References:

Database used

INTCAL09

References to INTCAL09 database

*Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,
Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192*

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-9.9;lab. mult=1)

Laboratory number: Beta-333932


Conventional radiocarbon age: 2630±30 BP

2 Sigma calibrated result: Cal BC 830 to 790 (Cal BP 2780 to 2740)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 800 (Cal BP 2750)

1 Sigma calibrated result: Cal BC 810 to 800 (Cal BP 2760 to 2740)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150, Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.9;lab. mult=1)

Laboratory number: Beta-333933


Conventional radiocarbon age: 2640±30 BP

**2 Sigma calibrated result: Cal BC 830 to 790 (Cal BP 2780 to 2740)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 800 (Cal BP 2750)

1 Sigma calibrated result: Cal BC 810 to 800 (Cal BP 2760 to 2750)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-8.7:lab. mult=1)

Laboratory number: Beta-333934

Conventional radiocarbon age: 2730±30 BP

2 Sigma calibrated result: Cal BC 920 to 810 (Cal BP 2870 to 2760)
(95% probability)

Intercept data


Intercepts of radiocarbon age
with calibration curve:

Cal BC 890 (Cal BP 2840) and

Cal BC 880 (Cal BP 2820) and

Cal BC 850 (Cal BP 2800)

1 Sigma calibrated result: Cal BC 900 to 830 (Cal BP 2850 to 2780)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-23:lab. mult=1)

Laboratory number: Beta-333935


Conventional radiocarbon age: 2380±30 BP

2 Sigma calibrated result: Cal BC 520 to 390 (Cal BP 2470 to 2340)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 400 (Cal BP 2360)

1 Sigma calibrated result: Cal BC 410 to 400 (Cal BP 2360 to 2350)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.2:lab. mult=1)

Laboratory number: Beta-333936
Conventional radiocarbon age: 2660±30 BP
2 Sigma calibrated result: Cal BC 840 to 800 (Cal BP 2790 to 2740)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 810 (Cal BP 2760)
1 Sigma calibrated result: Cal BC 830 to 800 (Cal BP 2780 to 2750)
(68% probability)



References:

Database used
INTCAL09
References to INTCAL09 database
Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,
Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192
Mathematics used for calibration scenario
A Simplified Approach to Calibrating C14 Dates
Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.2;lab. mult=1)

Laboratory number: Beta-333937


Conventional radiocarbon age: 2720±30 BP

**2 Sigma calibrated result: Cal BC 920 to 810 (Cal BP 2870 to 2760)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 840 (Cal BP 2790)

1 Sigma calibrated result: Cal BC 900 to 830 (Cal BP 2850 to 2780)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-7.9;lab. mult=1)

Laboratory number: Beta-333938
Conventional radiocarbon age: 2700±30 BP
2 Sigma calibrated result: Cal BC 900 to 800 (Cal BP 2850 to 2750)
 (95% probability)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 830 (Cal BP 2780)
 1 Sigma calibrated results: Cal BC 890 to 880 (Cal BP 2840 to 2820) and
 (68% probability) Cal BC 850 to 810 (Cal BP 2800 to 2760)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150, Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.8;lab. mult=1)

Laboratory number: Beta-333939


Conventional radiocarbon age: 2980±30 BP

**2 Sigma calibrated result: Cal BC 1310 to 1120 (Cal BP 3260 to 3070)
(95% probability)**

Intercept data

Intercepts of radiocarbon age
with calibration curve: Cal BC 1260 (Cal BP 3200) and
Cal BC 1240 (Cal BP 3190) and
Cal BC 1210 (Cal BP 3160)

1 Sigma calibrated results: Cal BC 1260 to 1190 (Cal BP 3210 to 3140) and
Cal BC 1180 to 1160 (Cal BP 3130 to 3110) and
Cal BC 1140 to 1130 (Cal BP 3090 to 3080)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,
Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates
Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.1:lab. mult=1)

Laboratory number: Beta-333940


Conventional radiocarbon age: 2920±30 BP

**2 Sigma calibrated results: Cal BC 1260 to 1240 (Cal BP 3200 to 3190) and
(95% probability) Cal BC 1210 to 1010 (Cal BP 3160 to 2960)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 1120 (Cal BP 3070)

**1 Sigma calibrated results: Cal BC 1190 to 1180 (Cal BP 3140 to 3130) and
(68% probability) Cal BC 1160 to 1140 (Cal BP 3110 to 3090) and
Cal BC 1130 to 1050 (Cal BP 3080 to 3000)**



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150, Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com



*Consistent Accuracy . . .
. . . Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President
Ronald Hatfield
Christopher Patrick
Deputy Directors

January 21, 2013

Dr. William H. Doelle/Jim Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Results For Samples LCAFN555, LCAFN1173, LCAFN1666, LCAFN1692, LCAFN1815, LCAFN1999, LCAFN2377, LCAFN5012, LCAFN5202, LCAFN5332, LCAFN5815, LCAFN6211

Dear Dr. Doelle & Mr. Vint:

Enclosed are the radiocarbon dating results for 12 samples recently sent to us. They each provided plenty of carbon for accurate measurements and all the analyses proceeded normally. As usual, the method of analysis is listed on the report with the results and calibration data is provided where applicable.

As always, no students or intern researchers who would necessarily be distracted with other obligations and priorities were used in the analyses. We analyzed them with the combined attention of our entire professional staff.

If you have specific questions about the analyses, please contact us. We are always available to answer your questions.

The cost of the analysis was charged to the VISA card provided. Thank you. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Digital signature on file


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 1/21/2013

Desert Archaeology, Incorporated

Material Received: 1/10/2013

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 339685 SAMPLE : LCAFN555 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 920 to 810 (Cal BP 2870 to 2760)	2450 +/- 30 BP	-8.5 o/oo	2720 +/- 30 BP
Beta - 339686 SAMPLE : LCAFN1173 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 780 to 520 (Cal BP 2730 to 2470)	2260 +/- 30 BP	-10.3 o/oo	2500 +/- 30 BP
Beta - 339687 SAMPLE : LCAFN1666 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 800 to 750 (Cal BP 2750 to 2700) AND Cal BC 680 to 670 (Cal BP 2630 to 2620) Cal BC 610 to 600 (Cal BP 2560 to 2550)	2330 +/- 30 BP	-10.8 o/oo	2560 +/- 30 BP
Beta - 339688 SAMPLE : LCAFN1692 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 800 to 750 (Cal BP 2750 to 2700) AND Cal BC 680 to 670 (Cal BP 2630 to 2620) Cal BC 610 to 600 (Cal BP 2560 to 2550)	2330 +/- 30 BP	-10.9 o/oo	2560 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the 14C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby 14C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured 13C/12C ratios (delta 13C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta 13C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta 13C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

**BETA ANALYTIC INC.**

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 1/21/2013

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 339689 SAMPLE : LCAFN1815 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 770 to 510 (Cal BP 2720 to 2460)	2470 +/- 30 BP	-23.5 o/oo	2490 +/- 30 BP
Beta - 339690 SAMPLE : LCAFN1999 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 810 to 770 (Cal BP 2760 to 2720)	2360 +/- 30 BP	-10.2 o/oo	2600 +/- 30 BP
Beta - 339691 SAMPLE : LCAFN2377 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 810 to 770 (Cal BP 2760 to 2720)	2350 +/- 30 BP	-9.8 o/oo	2600 +/- 30 BP
Beta - 339693 SAMPLE : LCAFN5012 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 780 to 520 (Cal BP 2730 to 2470)	2270 +/- 30 BP	-10.8 o/oo	2500 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX:305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 1/21/2013

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 339694 SAMPLE : LCAFN5202 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 780 to 520 (Cal BP 2730 to 2470)	2250 +/- 30 BP	-10.0 o/oo	2500 +/- 30 BP
Beta - 339695 SAMPLE : LCAFN5332 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 790 to 730 (Cal BP 2740 to 2680) AND Cal BC 690 to 660 (Cal BP 2640 to 2610) Cal BC 650 to 540 (Cal BP 2600 to 2490)	2280 +/- 30 BP	-9.8 o/oo	2530 +/- 30 BP
Beta - 339696 SAMPLE : LCAFN5815 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 760 to 680 (Cal BP 2710 to 2630) AND Cal BC 670 to 410 (Cal BP 2620 to 2360)	2220 +/- 30 BP	-11.1 o/oo	2450 +/- 30 BP
Beta - 339697 SAMPLE : LCAFN6211 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 910 to 810 (Cal BP 2860 to 2760)	2470 +/- 30 BP	-10.4 o/oo	2710 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the 14C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby 14C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured 13C/12C ratios (delta 13C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta 13C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta 13C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-8.5:lab. mult=1)

Laboratory number: Beta-339685


Conventional radiocarbon age: 2720±30 BP

**2 Sigma calibrated result: Cal BC 920 to 810 (Cal BP 2870 to 2760)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 840 (Cal BP 2790)

1 Sigma calibrated result: Cal BC 900 to 830 (Cal BP 2850 to 2780)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,
Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates
Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.3:lab. mult=1)

Laboratory number: Beta-339686
Conventional radiocarbon age: 2500±30 BP
2 Sigma calibrated result: Cal BC 780 to 520 (Cal BP 2730 to 2470)
(95% probability)


Intercept data

Intercepts of radiocarbon age
with calibration curve:

Cal BC 750 (Cal BP 2700) and
 Cal BC 680 (Cal BP 2630) and
 Cal BC 670 (Cal BP 2620) and
 Cal BC 610 (Cal BP 2560) and
 Cal BC 600 (Cal BP 2550)

1 Sigma calibrated results:
(68% probability)

Cal BC 760 to 730 (Cal BP 2720 to 2680) and
 Cal BC 690 to 680 (Cal BP 2640 to 2630) and
 Cal BC 670 to 660 (Cal BP 2620 to 2610) and
 Cal BC 650 to 540 (Cal BP 2600 to 2490)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150, Stuiver, et al., 1993, *Radiocarbon* 35(1):1-244, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.8;lab. mult=1)

Laboratory number: Beta-339687


Conventional radiocarbon age: 2560±30 BP

**2 Sigma calibrated results: Cal BC 800 to 750 (Cal BP 2750 to 2700) and
(95% probability) Cal BC 680 to 670 (Cal BP 2630 to 2620) and
Cal BC 610 to 600 (Cal BP 2560 to 2550)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 780 (Cal BP 2730)

1 Sigma calibrated results: Cal BC 790 to 760 (Cal BP 2740 to 2720) and
(68% probability) Cal BC 680 to 670 (Cal BP 2630 to 2620)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.9:lab. mult=1)

Laboratory number: Beta-339688
Conventional radiocarbon age: 2560±30 BP
2 Sigma calibrated results: Cal BC 800 to 750 (Cal BP 2750 to 2700) and
 (95% probability) Cal BC 680 to 670 (Cal BP 2630 to 2620) and
 Cal BC 610 to 600 (Cal BP 2560 to 2550)

Intercept data

Intercept of radiocarbon age
 with calibration curve: Cal BC 780 (Cal BP 2730)
1 Sigma calibrated results: Cal BC 790 to 760 (Cal BP 2740 to 2720) and
 (68% probability) Cal BC 680 to 670 (Cal BP 2630 to 2620)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et.al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et.al., 2009, Radiocarbon 51(4):1111-1150, Stuiver, et.al., 1993, Radiocarbon 35(1):137-189, Oeschger, et.al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-23.5;lab. mult=1)

Laboratory number: Beta-339689

Conventional radiocarbon age: 2490±30 BP

**2 Sigma calibrated result: Cal BC 770 to 510 (Cal BP 2720 to 2460)
(95% probability)**


Intercept data

Intercepts of radiocarbon age
with calibration curve:

Cal BC 750 (Cal BP 2700) and
Cal BC 690 (Cal BP 2640) and
Cal BC 660 (Cal BP 2620) and
Cal BC 640 (Cal BP 2590) and
Cal BC 590 (Cal BP 2540) and
Cal BC 580 (Cal BP 2530) and
Cal BC 570 (Cal BP 2520)

1 Sigma calibrated results:
(68% probability)

Cal BC 760 to 730 (Cal BP 2710 to 2680) and
Cal BC 690 to 680 (Cal BP 2640 to 2630) and
Cal BC 670 to 660 (Cal BP 2620 to 2610) and
Cal BC 650 to 540 (Cal BP 2600 to 2490)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.2:lab. mult=1)

Laboratory number: Beta-339690
Conventional radiocarbon age: 2600±30 BP
2 Sigma calibrated result: Cal BC 810 to 770 (Cal BP 2760 to 2720)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 800 (Cal BP 2740)
1 Sigma calibrated result: Cal BC 800 to 790 (Cal BP 2750 to 2740)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et.al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et.al., 2009, Radiocarbon 51(4):1111-1150, Stuiver, et.al., 1993, Radiocarbon 35(1):137-189, Oeschger, et.al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-9.8;lab. mult=1)

Laboratory number: Beta-339691


Conventional radiocarbon age: 2600±30 BP

**2 Sigma calibrated result: Cal BC 810 to 770 (Cal BP 2760 to 2720)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 800 (Cal BP 2740)

1 Sigma calibrated result: Cal BC 800 to 790 (Cal BP 2750 to 2740)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et.al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et.al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et.al., 1993, Radiocarbon 35(1):137-189, Oeschger, et.al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.8:lab.mult=1)

Laboratory number: Beta-339693
Conventional radiocarbon age: 2500±30 BP
2 Sigma calibrated result: Cal BC 780 to 520 (Cal BP 2730 to 2470)
(95% probability)


Intercept data

Intercepts of radiocarbon age
with calibration curve:

Cal BC 750 (Cal BP 2700) and
 Cal BC 680 (Cal BP 2630) and
 Cal BC 670 (Cal BP 2620) and
 Cal BC 610 (Cal BP 2560) and
 Cal BC 600 (Cal BP 2550)

1 Sigma calibrated results:
(68% probability)

Cal BC 760 to 730 (Cal BP 2720 to 2680) and
 Cal BC 690 to 680 (Cal BP 2640 to 2630) and
 Cal BC 670 to 660 (Cal BP 2620 to 2610) and
 Cal BC 650 to 540 (Cal BP 2600 to 2490)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et al., 1993, Radiocarbon 35(1):1-244, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10;lab. mult=1)

Laboratory number: Beta-339694

Conventional radiocarbon age: 2500±30 BP

**2 Sigma calibrated result: Cal BC 780 to 520 (Cal BP 2730 to 2470)
(95% probability)**


Intercept data

Intercepts of radiocarbon age
with calibration curve:

Cal BC 750 (Cal BP 2700) and
Cal BC 680 (Cal BP 2630) and
Cal BC 670 (Cal BP 2620) and
Cal BC 610 (Cal BP 2560) and
Cal BC 600 (Cal BP 2550)

1 Sigma calibrated results:
(68% probability)

Cal BC 760 to 730 (Cal BP 2720 to 2680) and
Cal BC 690 to 680 (Cal BP 2640 to 2630) and
Cal BC 670 to 660 (Cal BP 2620 to 2610) and
Cal BC 650 to 540 (Cal BP 2600 to 2490)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et al., 1993, Radiocarbon 35(1):1-244, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS


(Variables: C13/C12=-9.8:lab. mult=1)

Laboratory number: Beta-339695
Conventional radiocarbon age: 2530±30 BP
2 Sigma calibrated results: Cal BC 790 to 730 (Cal BP 2740 to 2680) and
 (95% probability) Cal BC 690 to 660 (Cal BP 2640 to 2610) and
 Cal BC 650 to 540 (Cal BP 2600 to 2490)

Intercept data

Intercepts of radiocarbon age
 with calibration curve: Cal BC 760 (Cal BP 2720) and
 Cal BC 680 (Cal BP 2630) and
 Cal BC 670 (Cal BP 2620)

1 Sigma calibrated results: Cal BC 780 to 750 (Cal BP 2730 to 2700) and
 (68% probability) Cal BC 680 to 670 (Cal BP 2630 to 2620) and
 Cal BC 610 to 600 (Cal BP 2560 to 2550)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.1:lab. mult=1)

Laboratory number: Beta-339696

Conventional radiocarbon age: 2450±30 BP


**2 Sigma calibrated results: Cal BC 760 to 680 (Cal BP 2710 to 2630) and
(95% probability) Cal BC 670 to 410 (Cal BP 2620 to 2360)**

Intercept data

Intercepts of radiocarbon age
with calibration curve:

Cal BC 700 (Cal BP 2650) and
Cal BC 700 (Cal BP 2650) and
Cal BC 540 (Cal BP 2490) and
Cal BC 530 (Cal BP 2480) and
Cal BC 520 (Cal BP 2470)

**1 Sigma calibrated results: Cal BC 740 to 690 (Cal BP 2690 to 2640) and
(68% probability) Cal BC 660 to 650 (Cal BP 2610 to 2600) and
Cal BC 550 to 480 (Cal BP 2500 to 2430) and
Cal BC 460 to 410 (Cal BP 2420 to 2360)**



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.4:lab. mult=1)

Laboratory number: Beta-339697
Conventional radiocarbon age: 2710±30 BP
2 Sigma calibrated result: Cal BC 910 to 810 (Cal BP 2860 to 2760)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 840 (Cal BP 2780)
1 Sigma calibrated result: Cal BC 900 to 820 (Cal BP 2850 to 2770)
(68% probability)



References:

Database used
 INTCAL09
References to INTCAL09 database
 Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,
 Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192
Mathematics used for calibration scenario
 A Simplified Approach to Calibrating C14 Dates
 Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com



*Consistent Accuracy . . .
... Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President

Ronald Hatfield
Christopher Patrick
Deputy Directors

March 15, 2013

Dr. William H. Doelle/Jim Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Results For Samples LCAFN2737, LCAFN2509, LCAFN7691, LCAFN10881, LCAFN12317

Dear Dr. Doelle & Mr. Vint:

Enclosed are the radiocarbon dating results for five samples recently sent to us. They each provided plenty of carbon for accurate measurements and all the analyses proceeded normally. As usual, the method of analysis is listed on the report with the results and calibration data is provided where applicable.

The web directory containing the table of results and PDF download also contains pictures including, most importantly the portion actually analyzed. These can be saved by opening them and right clicking. Also a cvs spreadsheet download option is available and a quality assurance report is posted for each set of results. This report contains expected vs measured values for 3-5 working standards analyzed simultaneously with your samples.

All results reported are accredited to ISO-17025 standards and all analyses were performed entirely here in our laboratories. Since Beta is not a teaching laboratory, only graduates trained in accordance with the strict protocols of the ISO-17025 program participated in the analyses. When interpreting the results, please consider any communications you may have had with us regarding the samples.

If you have specific questions about the analyses, please contact us. Your inquiries are always welcome.

Thank you for prepaying the analyses. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Digital signature on file


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 3/15/2013

Desert Archaeology, Incorporated

Material Received: 3/7/2013

Sample Data	Measured Radiocarbon Age	13C/12C Ratio	Conventional Radiocarbon Age(*)
Beta - 344167 SAMPLE : LCAFN2737 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 790 to 730 (Cal BP 2740 to 2680) AND Cal BC 690 to 660 (Cal BP 2640 to 2610) Cal BC 650 to 540 (Cal BP 2600 to 2490)	2270 +/- 30 BP	-9.7 o/oo	2520 +/- 30 BP
Beta - 344168 SAMPLE : LCAFN2509 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 800 to 740 (Cal BP 2740 to 2690) AND Cal BC 690 to 660 (Cal BP 2640 to 2610) Cal BC 650 to 550 (Cal BP 2600 to 2500)	2290 +/- 30 BP	-10.0 o/oo	2540 +/- 30 BP
Beta - 344169 SAMPLE : LCAFN7691 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 760 to 680 (Cal BP 2720 to 2630) AND Cal BC 670 to 410 (Cal BP 2620 to 2360)	2240 +/- 30 BP	-10.9 o/oo	2470 +/- 30 BP
Beta - 344170 SAMPLE : LCAFN10881 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 2570 to 2460 (Cal BP 4520 to 4420)	3750 +/- 30 BP	-10.6 o/oo	3990 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the 14C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby 14C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured 13C/12C ratios (delta 13C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta 13C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta 13C, the ratio and the Conventional Radiocarbon Age will be followed by "m". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

**BETA ANALYTIC INC.**

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 3/15/2013

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 344171 SAMPLE : LCAFN12317 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 3770 to 3650 (Cal BP 5720 to 5600)	4690 +/- 30 BP	-10.5 o/oo	4930 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-9.7:lab. mult=1)

Laboratory number: Beta-344167

Conventional radiocarbon age: 2520±30 BP

**2 Sigma calibrated results: Cal BC 790 to 730 (Cal BP 2740 to 2680) and
(95% probability) Cal BC 690 to 660 (Cal BP 2640 to 2610) and
Cal BC 650 to 540 (Cal BP 2600 to 2490)**


Intercept data

Intercepts of radiocarbon age
with calibration curve:

Cal BC 760 (Cal BP 2710) and
Cal BC 680 (Cal BP 2630) and
Cal BC 670 (Cal BP 2620)

1 Sigma calibrated results:
(68% probability)

Cal BC 770 to 750 (Cal BP 2720 to 2700) and
Cal BC 690 to 660 (Cal BP 2640 to 2620) and
Cal BC 640 to 590 (Cal BP 2590 to 2540) and
Cal BC 580 to 570 (Cal BP 2530 to 2520)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et al., 1993, Radiocarbon 35(1):137-189, Oeschger, et al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10:lab. mult=1)

Laboratory number: Beta-344168


Conventional radiocarbon age: 2540±30 BP

**2 Sigma calibrated results: Cal BC 800 to 740 (Cal BP 2740 to 2690) and
(95% probability) Cal BC 690 to 660 (Cal BP 2640 to 2610) and
Cal BC 650 to 550 (Cal BP 2600 to 2500)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 770 (Cal BP 2720)

1 Sigma calibrated results: Cal BC 790 to 760 (Cal BP 2740 to 2710) and
Cal BC 680 to 670 (Cal BP 2630 to 2620)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et.al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et.al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et.al., 1993, Radiocarbon 35(1):137-189, Oeschger, et.al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS


(Variables: C13/C12=-10.9:lab. mult=1)

Laboratory number: Beta-344169
Conventional radiocarbon age: 2470±30 BP
2 Sigma calibrated results: Cal BC 760 to 680 (Cal BP 2720 to 2630) and
(95% probability) Cal BC 670 to 410 (Cal BP 2620 to 2360)

Intercept data

Intercepts of radiocarbon age
with calibration curve: Cal BC 730 (Cal BP 2680) and
Cal BC 690 (Cal BP 2640) and
Cal BC 660 (Cal BP 2610) and
Cal BC 650 (Cal BP 2600) and
Cal BC 540 (Cal BP 2490)

1 Sigma calibrated results: Cal BC 750 to 680 (Cal BP 2700 to 2630) and
(68% probability) Cal BC 670 to 610 (Cal BP 2620 to 2560) and
Cal BC 600 to 520 (Cal BP 2550 to 2470)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et.al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et.al., 2009, Radiocarbon 51(4):1111-1150,

Stuiver, et.al., 1993, Radiocarbon 35(1):137-189, Oeschger, et.al., 1975, Tellus 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.6;lab. mult=1)

Laboratory number: **Beta-344170**


Conventional radiocarbon age: **3990±30 BP**

2 Sigma calibrated result: Cal BC 2570 to 2460 (Cal BP 4520 to 4420)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 2490 (Cal BP 4440)

1 Sigma calibrated results: Cal BC 2570 to 2520 (Cal BP 4520 to 4470) and
(68% probability) Cal BC 2500 to 2470 (Cal BP 4450 to 4420)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.5:lab. mult=1)

Laboratory number: Beta-344171
Conventional radiocarbon age: 4930±30 BP
2 Sigma calibrated result: Cal BC 3770 to 3650 (Cal BP 5720 to 5600)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 3700 (Cal BP 5650)
1 Sigma calibrated result: Cal BC 3710 to 3660 (Cal BP 5660 to 5610)
(68% probability)



References:

Database used
 INTCAL09
References to INTCAL09 database
 Heaton, et.al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et.al, 2009, Radiocarbon 51(4):1111-1150,
 Stuiver, et.al, 1993, Radiocarbon 35(1):137-189, Oeschger, et.al., 1975, Tellus 27:168-192
Mathematics used for calibration scenario
 A Simplified Approach to Calibrating C14 Dates
 Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com



*Consistent Accuracy . . .
... Delivered On-time*

Beta Analytic Inc.
4985 SW 74 Court
Miami, Florida 33155 USA
Tel: 305 667 5167
Fax: 305 663 0964
Beta@radiocarbon.com
www.radiocarbon.com

Darden Hood
President

Ronald Hatfield
Christopher Patrick
Deputy Directors

September 6, 2013

Dr. William H. Doelle/ Mr. Jim Vint
Desert Archaeology, Incorporated
3975 North Tucson Boulevard
Tucson, AZ 85716
USA

RE: Radiocarbon Dating Results For Samples LCAFN16032, LCAFN16060, LCAFN16240,
LCAFN16252

Dear Dr. Doelle & Mr. Vint:

Enclosed are the radiocarbon dating results for four samples recently sent to us. They each provided plenty of carbon for accurate measurements and all the analyses proceeded normally. As usual, the method of analysis is listed on the report with the results and calibration data is provided where applicable.

The web directory containing the table of results and PDF download also contains pictures including, most importantly the portion actually analyzed. These can be saved by opening them and right clicking. Also a cvs spreadsheet download option is available and a quality assurance report is posted for each set of results. This report contains expected versus measured values for 3-5 working standards analyzed simultaneously with your samples.

All results reported are accredited to ISO-17025 standards and all analyses were performed entirely here in our laboratories. Since Beta is not a teaching laboratory, only graduates trained in accordance with the strict protocols of the ISO-17025 program participated in the analyses. When interpreting the results, please consider any communications you may have had with us regarding the samples.

If you have specific questions about the analyses, please contact us. Your inquiries are always welcome.

Thank you for prepaying the analyses. As always, if you have any questions or would like to discuss the results, don't hesitate to contact me.

Sincerely,

Digital signature on file


BETA ANALYTIC INC.

DR. M.A. TAMERS and MR. D.G. HOOD

 4985 S.W. 74 COURT
 MIAMI, FLORIDA, USA 33155
 PH: 305-667-5167 FAX: 305-663-0964
 beta@radiocarbon.com

REPORT OF RADIOCARBON DATING ANALYSES

Dr. William H. Doelle/Jim Vint

Report Date: 9/6/2013

Desert Archaeology, Incorporated

Material Received: 8/28/2013

Sample Data	Measured Radiocarbon Age	¹³ C/ ¹² C Ratio	Conventional Radiocarbon Age(*)
Beta - 358013 SAMPLE : LCAFN16032 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 520 to 390 (Cal BP 2470 to 2340)	2130 +/- 30 BP	-9.5 o/oo	2380 +/- 30 BP
Beta - 358014 SAMPLE : LCAFN16060 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 790 to 730 (Cal BP 2740 to 2680) AND Cal BC 690 to 660 (Cal BP 2640 to 2610) Cal BC 650 to 540 (Cal BP 2600 to 2490)	2300 +/- 30 BP	-11.1 o/oo	2530 +/- 30 BP
Beta - 358015 SAMPLE : LCAFN16240 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 840 to 790 (Cal BP 2780 to 2740)	2410 +/- 30 BP	-10.3 o/oo	2650 +/- 30 BP
Beta - 358016 SAMPLE : LCAFN16252 ANALYSIS : AMS-Standard delivery MATERIAL/PRETREATMENT : (charred material): acid/alkali/acid 2 SIGMA CALIBRATION : Cal BC 810 to 770 (Cal BP 2760 to 2720)	2550 +/- 30 BP	-22.1 o/oo	2600 +/- 30 BP

Dates are reported as RCYBP (radiocarbon years before present, "present" = AD 1950). By international convention, the modern reference standard was 95% the ¹⁴C activity of the National Institute of Standards and Technology (NIST) Oxalic Acid (SRM 4990C) and calculated using the Libby ¹⁴C half-life (5568 years). Quoted errors represent 1 relative standard deviation statistics (68% probability) counting errors based on the combined measurements of the sample, background, and modern reference standards. Measured ¹³C/¹²C ratios (delta ¹³C) were calculated relative to the PDB-1 standard.

The Conventional Radiocarbon Age represents the Measured Radiocarbon Age corrected for isotopic fractionation, calculated using the delta ¹³C. On rare occasion where the Conventional Radiocarbon Age was calculated using an assumed delta ¹³C, the ratio and the Conventional Radiocarbon Age will be followed by "**". The Conventional Radiocarbon Age is not calendar calibrated. When available, the Calendar Calibrated result is calculated from the Conventional Radiocarbon Age and is listed as the "Two Sigma Calibrated Result" for each sample.

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-9.5:lab. mult=1)

Laboratory number: Beta-358013


Conventional radiocarbon age: 2380±30 BP

**2 Sigma calibrated result: Cal BC 520 to 390 (Cal BP 2470 to 2340)
(95% probability)**

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 400 (Cal BP 2360)

1 Sigma calibrated result: Cal BC 410 to 400 (Cal BP 2360 to 2350)
(68% probability)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et.al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et.al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et.al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et.al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-11.1:lab. mult=1)

Laboratory number: Beta-358014
Conventional radiocarbon age: 2530±30 BP
2 Sigma calibrated results: Cal BC 790 to 730 (Cal BP 2740 to 2680) and
 (95% probability) Cal BC 690 to 660 (Cal BP 2640 to 2610) and
 Cal BC 650 to 540 (Cal BP 2600 to 2490)


Intercept data

Intercepts of radiocarbon age
 with calibration curve:

Cal BC 760 (Cal BP 2720) and
 Cal BC 680 (Cal BP 2630) and
 Cal BC 670 (Cal BP 2620)

1 Sigma calibrated results:
 (68% probability)

Cal BC 780 to 750 (Cal BP 2730 to 2700) and
 Cal BC 680 to 670 (Cal BP 2630 to 2620) and
 Cal BC 610 to 600 (Cal BP 2560 to 2550)



References:

Database used

INTCAL09

References to INTCAL09 database

Heaton, et al., 2009, *Radiocarbon* 51(4):1151-1164, Reimer, et al., 2009, *Radiocarbon* 51(4):1111-1150,

Stuiver, et al., 1993, *Radiocarbon* 35(1):137-189, Oeschger, et al., 1975, *Tellus* 27:168-192

Mathematics used for calibration scenario

A Simplified Approach to Calibrating C14 Dates

Talma, A. S., Vogel, J. C., 1993, *Radiocarbon* 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-10.3;lab. mult=1)

Laboratory number: Beta-358015
Conventional radiocarbon age: 2650±30 BP
2 Sigma calibrated result: Cal BC 840 to 790 (Cal BP 2780 to 2740)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 810 (Cal BP 2760)
1 Sigma calibrated result: Cal BC 820 to 800 (Cal BP 2770 to 2750)
(68% probability)



References:

- Database used*
INTCAL09
- References to INTCAL09 database*
Heaton, et.al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et.al., 2009, Radiocarbon 51(4):1111-1150,
Stuiver, et.al., 1993, Radiocarbon 35(1):137-189, Oeschger, et.al., 1975, Tellus 27:168-192
- Mathematics used for calibration scenario*
A Simplified Approach to Calibrating C14 Dates
Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com


CALIBRATION OF RADIOCARBON AGE TO CALENDAR YEARS

(Variables: C13/C12=-22.1:lab. mult=1)

Laboratory number: Beta-358016
Conventional radiocarbon age: 2600±30 BP
2 Sigma calibrated result: Cal BC 810 to 770 (Cal BP 2760 to 2720)
(95% probability)

Intercept data

Intercept of radiocarbon age
with calibration curve: Cal BC 800 (Cal BP 2740)
1 Sigma calibrated result: Cal BC 800 to 790 (Cal BP 2750 to 2740)
(68% probability)



References:

Database used
 INTCAL09
References to INTCAL09 database
 Heaton, et.al., 2009, Radiocarbon 51(4):1151-1164, Reimer, et.al., 2009, Radiocarbon 51(4):1111-1150,
 Stuiver, et.al., 1993, Radiocarbon 35(1):137-189, Oeschger, et.al., 1975, Tellus 27:168-192
Mathematics used for calibration scenario
 A Simplified Approach to Calibrating C14 Dates
 Talma, A. S., Vogel, J. C., 1993, Radiocarbon 35(2):317-322

Beta Analytic Radiocarbon Dating Laboratory

4985 S.W. 74th Court, Miami, Florida 33155 • Tel: (305)667-5167 • Fax: (305)663-0964 • E-Mail: beta@radiocarbon.com

